

SOURCE APPORTIONMENT OF CARBONACEOUS AEROSOLS WITH HIGH TIME RESOLUTION

J. T. Coutinho¹, M. Almeida-Silva¹, C. Galinha¹, N. Canha¹, A. Alaustey², M. Rigler³, G. Močnik⁴, S. Marta Almeida¹

¹C2TN, IST, Universidade de Lisboa, EN 10 ao km 139.7, 2695-066, Bobadela, Portugal

²IDÆA, Spanish National Research Council, C. Jordi Girona 18-26, 08034, Barcelona, Spain

³Aerosol d.o.o., Kamniška ulica 39a, 1000, Ljubljana, Slovenia

⁴Jozef Stefan Institute, Jamova cesta 39, 1000, Ljubljana, Slovenia

Keywords: black carbon, Aethalometer, Total Carbon Analyzer, thermal protocol

Presenting author email: coutinho.joana@ctn.tecnico.ulisboa.pt

Chemical composition of aerosols is characterised by a large spatio-temporal heterogeneity, being carbonaceous aerosols the major components of the submicron fraction of atmospheric particulate matter (PM) (Mohr, 2011). They are emitted by different sources that exert a negative impact on human health, and also affect the climate and the environment. (UNEP-CCAC, 2014). The components of carbonaceous PM (total carbon, TC) are organic carbon (OC), elemental carbon (EC) and inorganic carbon (IC). When EC is measured using optical methods relying on its strongly light absorbing character it is called black carbon (BC), which is emitted during the incomplete combustion of fossil fuels, biofuels, and biomass burning and absorbs at all wavelengths of solar radiation (Becerril-Valle, 2017).

It is thus of paramount importance to determine the chemical composition of submicron PM at high time resolution, providing the necessary information for accurate source apportionment. With this purpose a campaign was conducted in the urban centre of Moscavide (North of Lisbon, Portugal). Different fractions of air PM were collected and online BC (Aethalometer AE33) and TC measurements (Total Carbon Analyzer TCA08) were performed, simultaneously. The sampled filters were analysed by gravimetry and thermo-optical analysis for the measurement of OC/EC (using two different thermal protocols). The Aethalometer Model (Sandradewi, 2008) was applied for the BC source apportionment due to fossil fuel (BC_{ff}) and biomass burning (BC_{bb}) contributions (Fig. 1). The recently developed TC-BC online method, which combines an optical method for measuring BC by the AE33 (Drinovec, 2015) and a thermal method for TC determination by the TCA08, was used for source apportionment of carbonaceous aerosols with high time resolution. This method determines equivalent OC fraction (eOC) of carbonaceous aerosols that is the difference between TC and

EC (inferred from BC), at high time resolution, $eOC = TC - b \cdot BC$. The determined proportionality parameter b is region/site specific and depends to a large extent on a thermal protocol used to determine the EC fraction with the conventional OC/EC method.

The combination of the data generated by the online equipments with data from the analysis of the offline filters allowed us to obtain the parameters for the EUSAAR2 protocol.




Fig. 1. Source apportionment of BC using $\alpha_{ff} = 1.0$ and $\alpha_{bb} = 2.0$ (green: BC_{ff}, orange: BC_{bb}).

This work was supported by Life Program - Life Index Air project, Interreg Med - REMEDIO project and Cost Action - CA16109 COLOSSAL.

Becerril-Valle, M., Coz, E., Prévôt, A.S.H., Mocnik, G., Pandis, S.N., de la Campa, A.M. S., Alastuey, A., Díaz, E., Pérez, R.M., Artíñano, B. (2017). Atmos. Environ, 169, 36-53.

Drinovec, L., Mocnik, G., Zotter, P., Prévôt, A.S.H., et al. (2015). Atmos. Meas. Tech., 8, 1965–1979.

Mohr, C., Richter, R., DeCarlo, P. F., Prévôt, A. S. H., and Baltensperger, U. (2011). Atmos. Chem. Phys., 11, 7465-7482.

Sandradewi, J., Prévôt, A.S.H., Szidat, S., Perron, N., Alfarra, M.A., Lanz, V.A., Weingartner, E., Baltensperger, U. (2008). Environ. Sci. Technol., 42, 3316-3323.

UNEP-CCAC (2014). Time to act to reduce short-lived climate pollutants.